

**CERTIFICATE OF COMPLIANCE**
***CERTIFICAT DE CONFORMITE***

Product <i>Produit</i>	HSV straw x 4 Paillette VHS x 4
Product code <i>Code Article</i>	025246
Batch# <i>Lot N°</i>	CBS 016645
Expiration date <i>Date de péremption</i>	2022-09

We certify that the following controlled goods are in compliance with our internal specifications.

*Nous certifions que les produits contrôlés suivants sont en conformité avec nos spécifications internes.*

Have been sterilized by irradiation, and are in conformance with ISO 13485 requirements for Quality Management system of medical devices and ISO 11137 requirements for process control and routine monitoring of radiation sterilization of health care products.

*Ont été stérilisés par irradiation, et sont en conformité avec les exigences de l'ISO 13485 pour le système de management de la qualité de dispositifs médicaux, et l'ISO 11137 pour les exigences et le contrôle de routine des stérilisations par irradiation des dispositifs médicaux.*

Specifications <i>Spécifications</i>	Acceptance criteria <i>Critères d'acceptation</i>	Results <i>Résultats</i>
MEA Test	Min 80% development to blastocyst at 96h	Compliant
Endotoxin Test	<1UI/device	Compliant

Date : 24/10/2019

M. HIRIGARAY  
Qualified Person

